

HyperArc Radiosurgery Bibliography*

- Ohira S, Sagawa T, Ueda Y, Inui S, Masaoka A, Akino Y, Mizuno H, Miyazaki M, Koizumi M, Teshima T. Effect of collimator angle on HyperArc™ stereotactic radiosurgery planning for single and multiple brain metastases. *Med Dosim*. 2019 Aug 1. Osaka International Cancer Institute, Osaka, Japan [Epub ahead of print]
- Kadoya N, Abe Y, Kajikawa T, Ito K, Yamamoto T, Umezawa R, Chiba T, Katsuta Y, Takayama Y, Kato T, Kikuchi Y, Jingu K. Automated noncoplanar treatment planning strategy in stereotactic radiosurgery of multiple cranial metastases: HyperArc™ and CyberKnife® dose distributions. *Med Dosim*. 2019 Winter;44(4):394-400. Tohoku University Graduate School of Medicine, Sendai, Japan
- Sagawa T, Ohira S, Ueda Y, Akino Y, Mizuno H, Matsumoto M, Miyazaki M, Teshima T, Koizumi M. Dosimetric effect of rotational setup errors in stereotactic radiosurgery with HyperArc™ for single and multiple brain metastases. *J Appl Clin Med Phys*. 2019 Oct;20(10):84-91. Osaka International Cancer Institute, Osaka, Japan
- Ueda Y, Ohira S, Yamazaki H, Mabuchi N, Higashinaka N, Miyazaki M, Teshima T. Dosimetric performance of two linear accelerator-based radiosurgery systems to treat single and multiple brain metastases. *Br J Radiol*. 2019 Aug;92(1100):20190004. Osaka International Cancer Institute, Osaka, Japan
- Hartgerink D, Swinnen A, Roberge D, Nichol A, Zygmanski P, Yin FF, Deblois F, Hurkmans C, Ong CL, Bruynzeel A, Aizer A, Fiveash J, Kirckpatrick J, Guckenberger M, Andratschke N, de Ruyscher D, Popple R, Zindler J. LINAC based stereotactic radiosurgery for multiple brain metastases: guidance for clinical implementation. *Acta Oncol*. 2019 Jul 1:1-8. Multinstitution led by Maastricht, Maastricht, The Netherlands
- Vergalasova I, Liu H, Alonso-Basanta M, Dong L, Li J, Nie K, Shi W, Teo BK, Yu Y, Yue NJ, Zou W, Li T. Multi-Institutional Dosimetric Evaluation of Modern Day Stereotactic Radiosurgery (SRS) Treatment Options for Multiple Brain Metastases. *Front Oncol*. 2019 Jun 7;9:483. Rutgers University, New Brunswick, NJ
- Ruggieri R, Naccarato S, Mazzola R, Ricchetti F, Corradini S, Fiorentino A, Alongi F. Linac-based radiosurgery for multiple brain metastases: Comparison between two mono-isocenter techniques with multiple non-coplanar arcs. *Radiother Oncol*. 2019 Mar;132:70-78. Ospedale "Sacro Cuore", Negrar, Italy
- Alongi F, Fiorentino A, Gregucci F, Corradini S, Gaj-Levra N, Romano L, Rigo M, Ricchetti F, Beltramello A, Lunardi G, Mazzola R, Ruggieri R. First experience and clinical results using a new non-coplanar mono-isocenter technique (HyperArc™) for Linac-based VMAT radiosurgery in brain metastases. *J Cancer Res Clin Oncol*. 2019 Jan;145(1):193-200. Ospedale "Sacro Cuore", Negrar, Italy
- Alongi F, Fiorentino A, Ruggieri R, Ricchetti F, Kupelian P. Cost-effectiveness of Linac-based single-isocenter non-coplanar technique (HyperArc™) for brain metastases radiosurgery. *Clin Exp Metastasis*. 2018 Oct;35(7):601-603. Ospedale "Sacro Cuore", Negrar, Italy
- Slosarek K, Bekman B, Wendykier J, Grządziel A, Fogliata A, Cozzi L. In silico assessment of the dosimetric quality of a novel, automated radiation treatment planning strategy for linac-based radiosurgery of multiple brain metastases and a comparison with robotic methods. *Radiat Oncol*. 2018 Mar 15;13(1):41. Maria Skłodowska Curie Memorial Cancer Center and Institute of Oncology, Gliwice, Poland
- Ruggieri R, Naccarato S, Mazzola R, Ricchetti F, Corradini S, Fiorentino A, Alongi F. Linac-based VMAT radiosurgery for multiple brain lesions: comparison between a conventional multi-isocenter approach and a new dedicated mono-isocenter technique. *Radiat Oncol*. 2018 Mar 5;13(1):38. Ospedale "Sacro Cuore", Negrar, Italy
- Ohira S, Ueda Y, Akino Y, Hashimoto M, Masaoka A, Hirata T, Miyazaki M, Koizumi M, Teshima T. HyperArc™ VMAT planning for single and multiple brain metastases stereotactic radiosurgery: a new treatment planning approach. *Radiat Oncol*. 2018 Jan 29;13(1):13. Osaka International Cancer Institute, Osaka, Japan

* This bibliography is a comprehensive selection of articles but is not necessarily an exhaustive list of literature pertaining to HyperArc™ radiosurgery.

Intended Use Summary

Varian Medical Systems' linear accelerators are intended to provide stereotactic radiosurgery and precision radiotherapy for lesions, tumors, and conditions anywhere in the body where radiation treatment is indicated.

Safety Statement

Radiation treatments may cause side effects that can vary depending on the part of the body being treated. The most frequent ones are typically temporary and may include, but are not limited to, irritation to the respiratory, digestive, urinary or reproductive systems, fatigue, nausea, skin irritation, and hair loss. In some patients, they can be severe. Treatment sessions may vary in complexity and time. Radiation treatment is not appropriate for all cancers.

varian

varian.com

USA, Corporate Headquarters and Manufacturer

Varian Medical Systems, Inc.
3100 Hansen Way
Palo Alto, CA 94304
Tel: 650.424.5700
800.544.4636

Headquarters Europe, Eastern Europe, Middle & Near East, India, Africa

Varian Medical Systems
International AG
Steinhausen, Switzerland
Tel: 41.41.749.8844

Asia Pacific Headquarters

Varian Medical Systems
Pacific, Inc.
Kowloon, Hong Kong
Tel: 852.2724.2836

Australasian Headquarters

Varian Medical Systems
Australasia Pty Ltd.
Sydney, Australia
Tel: 61.2.9485.0111

Latin American Headquarter

Varian Medical Systems
Brasil Ltda.
São Paulo, Brazil
Tel: 55.11.3457.2655

© 2019 Varian Medical Systems, Inc. All rights reserved. Varian and Varian Medical Systems are registered trademarks, and HyperArc is a trademark of Varian Medical Systems, Inc. All other trademarks are the property of their respective owners.